

Sicherheitsdatenblatt auf Anfrage
für berufsmäßige Verwender
erhältlich.

Safety data sheet available for
professional user on request.

Fiche de données de sécurité
disponible sur demande pour les
professionnels.

For laboratory use only

+2°C
+8°C

YYY-MM-DD

Lot XXXXX



REF D-0905-025-TF 01

**TissueFISH Pretreatment
Kit Part 2 of 2
Protease Stock Solution**



For Professional Use Only

Further information available at www.metasystems-probes.com

Product	Order No.	Pack Size
TissueFISH Pretreatment Kit Part 2 of 2 Protease Stock Solution	D-0905-025-TF	500µl

Description

Protease Stock Solution is part of the MetaSystems TissueFISH Pretreatment Kit (D-0905-025-TF). This Kit contains two parts, which have to be stored at different conditions. The MetaSystems TissueFISH Pretreatment Kit contains reagents and solutions to preprocess paraffin-embedded tissue sections for FISH assays. Reagents supplied are sufficient to run 20-30 Slides if processed in batches of 4-6 slides per run.

Materials Provided

Part 2 of MetaSystems TissueFISH Pretreatment Kit (Packing Lot: XXXXX) contains 500µl Protease Stock Solution with Lot: XXXXX-A

Intended Use

MetaSystems TissueFISH Pretreatment Kit is intended to be used for pretreatment of formalin-fixed, paraffin- embedded (FFPE) tissue sections.

Safety Instructions

	ATTENTION: Good Laboratory Practice! Use in accordance with the principles of good laboratory practice.
	ATTENTION: Waste Disposal! All hazardous materials should be disposed of according to your institution's guidelines for hazardous waste disposal.

Storage and Handling

Protease Stock Solution should be stored at +2 - +8°C.

Procedure

Follow the "MetaSystems TissueFISH Pretreatment Protocol for Paraffin-Embedded Tissue sections".







Customer Support

Please contact MetaSystems Probes GmbH in Germany or MetaSystems Group, Inc. in the USA by telephone or e-mail (contact details, see below). MetaSystems Probes disclaims any proprietary interest in the marks and names of others.



MetaSystems Probes GmbH
1. Industriestraße 7
D-68804 Altlußheim
Germany
Tel.: +49 (0)6205 292760
Fax: +49 (0)6205 2927629
email: info@metasystems-probes.com
URL: www.metasystems-probes.com

Symbols Used

Symbol	Description
	Detailed address and contact details of MetaSystems (manufacturer).
	All warnings are marked by warning triangle with exclamation mark. Depending on their character they are supplemented with the words ATTENTION or CAUTION.
	Order number of the respective probe
	Upper limit of the storage temperature
	Lot number
	Expiry date

Revision: Revision: TissueFISH-Part2-RevE210323-220623